# **Crush Tester Accessory**

### Corrugated fiberboard Edge Crush Test/adhesive sample cutter

### **Technical data:**

1, Sample size: 25mm×100mm (ECT), 25mm×80mm (adhesive)

Result error: ±0.5mm
 Maximum length: 220mm
 Maximum thickness: 18mm

5, Dimension (lxwxh): 430mmx380mmx200mm

6, weight: 22kg

### **Edge Crush Test metal block**

It is two blocks to fixate the specimen

Size: 10×2×2(cm) Weight: 0.2kg

# Corrugated fiberboard Flat Crush Te

#### Technical data:

1, Sample area: (two types): 64.5cm<sup>2</sup> 32.2cm<sup>2</sup>

2, Result error: ±0.5cm<sup>2</sup>
3, Sample thickness: ≤7mm

4, weight: 2.5kg

### **Ring Crush Test sample cutter**

### Technical data:

1, Sample length: 152 mm; Sample width: 12.7 mm;

Two length sides parallelism: <0.015mm

2, Sample thickness: (0.1~1.0)mm

3, Dimension (lxwxh): 570x190x368(mm)

4, weight: 19.5kg

### **RCT** central set

#### Technical data:

- 1. Ring groove width: 0.15mm ~ 1.00mm, 13 kinds of specifications
- 2. Dimensions (I  $\times$  w  $\times$  h): 440  $\times$  190  $\times$  48 mm

3. Weight: 2 kg

# Corrugated fiberboard Pin Adhesive Strength Test clamp (PAT)

#### Technical data:

- 1, Needle diameter: φ3mm (A flute), φ2mm (B, C flute) φ1mm (E flute)
- 2, Result error: assembled <1/100; pressed (300N) <1/200
- 3, Maximum dimension (lxwxh)): 100x50x45(mm)
- 4, weight: 0.6kg









